ProKnow Plan Study Review

Study Name: Participant Name: Participant Email: Participant Role: Participant Institution: Created: **TPS Manufacturer: TPS Model:** Number of Treatment Beams: Radiation Type(s): Delivery Type(s):

2023 AAMD Plan Study, Phase 1: Initial Head/Neck Plan

Lesley Rosa

lesley.rosa@varian.com

dosimetrist

Varian

Mon Jun 19 2023 23:19:38 GMT+0000

Varian Medical Systems

Eclipse

19

Photon


IMRT (Dynamic)

ProKnow

METRIC	RESULT	MIN REQ			IDEAL	PERFORMANCE (PTS)	WEIGHT
Volume (%) of the PTV63 covered by 63 (Gy)	98.794	90 🗸	2.5p	5p 7.5p 95 97	188 100	GOOD (9.00)	10.00
Volume (%) of the PTV60 covered by 60 (Gy)	99.021	90 🗸	2.5p	5p 7.5p 95 97	108 100	GOOD (9.18)	10.00
Volume (%) of the PTV57 covered by 57 (Gy)	99.420	90 🧹	3.13p 90	6.25p 9.38p 95 97	$\frac{12.5p}{100}$ 100	GOOD (11.90)	12.50
Volume (%) of the PTV54 covered by 54 (Gy)	99.121	90 🧹	3.13p 90	6.25p 9.38p 95 97	$\frac{12.5p}{100}$ 100	GOOD (11.58)	12.50
Dose (Gy) covering 99 (%) of the CTV63	63.214	55 😽	2.5p 55	5p 7.5 59 61	^{5p} ^{10p} 63	V IDEAL (10.00)	10.00
Dose (Gy) covering 99 (%) of the CTV60	60.521	56 🧹	2.5p 56	5p 7.5 58 59	^{5p} ^{10p} 60	V IDEAL (10.00)	10.00
Dose (Gy) covering 99 (%) of the CTV57	57.606	54 🗸	2.5p 55	7.5p 56	10p 57	V IDEAL (10.00)	10.00
Dose (Gy) covering 99 (%) of the CTV54	54.324	51 😽	2.5p 52	7.5p 53	^{10p} 54	V IDEAL (10.00)	10.00
High dose volume of regret (cc) [67 (Gy), CTV63]	0.093	1	7 1 ² P	3p 0.5	4p 5p 0.03	GOOD (4.63)	5.00
Dose (Gy) covering 0.03 (cc) of the SpinalCord	33.114	48 😽	2 8 38 48		<u></u> 38 30	GOOD (4.69)	5.00
Dose (Gy) covering 0.03 (cc) of the Brainstem	31.026	52 😽	7 <u>3</u> <u>5</u>	48 48	<u></u> 38 30	GOOD (4.90)	5.00
Volume (%) of the Parotid_L covered by 30 (Gy)	41.757	65 🥳	2 8 39	4p 45	§g 35	GOOD (4.32)	5.00
Volume (%) of the Parotid_R covered by 30 (Gy)	16.002	50 😽	2 ≩6 3§		48 IS 15	GOOD (4.80)	5.00
Mean dose (Gy) to the Parotid_L	27.923	40 😽	2 46 39	4B 30	26	GOOD (4.52)	5.00
Mean dose (Gy) to the Parotid_R	20.399	30 😽	7 38 38	4p 23	<u></u> ₹8 20	GOOD (4.87)	5.00
Mean dose (Gy) to the oral avoid	31.348	45 😽	2 48 48	4p 35	<u></u> 38 30	GOOD (4.73)	5.00
Mean dose (Gy) to the esophagus	17.274	30 😽	7 38 39		28 <u>18</u> 18	V IDEAL (5.00)	5.00
Dose (Gy) covering 0.03 (cc) of the pharyngeal constrictors	57.691	65 🧹	? 25 35	4p 45	<u></u> 3₿ 35	MARGINAL (2.73)	5.00
Mean dose (Gy) to the pharyngeal constrictors	37.156	50 😽	7 38 38		48 ⁵⁹ 37.5 37.5	V IDEAL (5.00)	5.00
Dose (Gy) covering 0.03 (cc) of the BP Avoid	64.191	68 🧹	? 28	3p 4p 63 62	₹8 60	MARGINAL (2.76)	5.00
Global maximum dose (Gy)	68.277	71 😽	7 38	4p 67	習 65	ACCEPTABLE (3.36)	5.00
[CRITICAL] Number of unique isocenters	1.000	≤1 🗸	7 -10p ≤	1	Op	IDEAL (0.00)	0.00
[CRITICAL] Number of unique couch angles	1.000	≤1 🗸	7 -10p ≤	1	Op	IDEAL (0.00)	0.00
Cumulative meterset over all treatment beams	3587.900						
TOTALS		23 (of 23)			8 (of 23)	137.97	150.00




Differential OMH





Differential OMH (Weighted)





----- PATIENT DATA SUMMARY ------

RT PLAN: Yes Patient Name: eByLuKOZoWxBUrIW Patient ID: vm3R2YZueEZB3INMHzS2pKFi6 Plan Name: ---Plan Label: 19f v18o18cI3 Study ID: ---Patient Position: HeadFirstSupine Manufacturer: Varian Medical Systems Model Name: ARIA RadOnc Total Beams: 20 [19 treatment, 1 setup, 0 port, 0 other] Total Control Points: 3156 Number of Fraction Groups: 1 [30 Fx]

RT DOSE: Yes Patient Name: eByLuKOZoWxBUrIW Patient ID: vm3R2YZueEZB3INMHzS2pKFi6 Study ID: ---Patient Position (Derived): HeadFirstSupine Patient Position (Requested): HeadFirstSupine Manufacturer: Varian Medical Systems Model Name: ARIA RadOnc Global Max Dose (Gy): 68.27741 X (mm): -247.004 to 247.996 step 2.5 Y (mm): -153.0 to 212.0 step 2.5 Z (mm): -170.678 to 114.322 step 2.5 DICOM Origin (mm): (0.00, 0.00, 0.00)

RT STRUCTURE SET: Yes Patient Name: AAMD^2023PlanStudy Patient ID: AAMD2023PLANSTUDY Structure Set Label: RTstruct Study ID: ORIGINAL Patient Position (Requested): HeadFirstSupine Slice Spacing: Uniform spacing of 2.5 mm Number of Structures: 34 [34 contour-based, 0 points]

IMAGE SET: Yes Patient Name: AAMD^2023PlanStudy Patient ID: AAMD2023PLANSTUDY Study ID: ORIGINAL Patient Position (Derived): HeadFirstSupine Patient Position (Requested): HeadFirstSupine Modality: ComputedTomography Axial Slices: 272 [2.5 mm spacing]



----- DICOM Alerts ------

The following messages report any inconsistencies found in your DICOM data associations. Please analyze each message to determine if they are important (e.g.

Inconsistency in Patient ID: vm3R2YZueEZB3INMHzS2pKFi6 [RT Plan] AAMD2023PLANSTUDY [RT Structure Set] vm3R2YZueEZB3INMHzS2pKFi6 [RT Dose] AAMD2023PLANSTUDY [CT Imageset]

Inconsistency in Patient Name: eByLuKOZoWxBUrIW [RT Plan] AAMD^2023PlanStudy [RT Structure Set] eByLuKOZoWxBUrIW [RT Dose] AAMD^2023PlanStudy [CT Imageset]

Inconsistency in Study ID: <Null> [RT Plan] ORIGINAL [RT Structure Set] <Null> [RT Dose] ORIGINAL [CT Imageset]

Inconsistency in Plan <-> Structure Set: The current RT Plan does not reference the SOP Instance UID of the current RT Structure Set.



BEAM [#] NAME	MACHINE
[01] Field 1	RDS3 Ti kV
[02] Field 2	RDS3 Ti kV
[03] Field 3	RDS3 Ti kV
[04] Field 4	RDS3 Ti kV
[05] Field 5	RDS3 Ti kV
[06] Field 6	RDS3 Ti kV
[07] Field 7	RDS3 Ti kV
[08] Field 8	RDS3 Ti kV
[09] Field 9	RDS3 Ti kV
[10] Field 10	RDS3 Ti kV
[11] Field 11	RDS3 Ti kV
[12] Field 12	RDS3 Ti kV
[13] Field 13	RDS3 Ti kV
[14] Field 14	RDS3 Ti kV
[15] Field 15	RDS3 Ti kV
[16] Field 16	RDS3 Ti kV
[17] Field 17	RDS3 Ti kV
[18] Field 18	RDS3 Ti kV
[19] Field 19	RDS3 Ti kV

BEAM [#] NAME

[01] Field 1	
[02] Field 2	
[03] Field 3	
[04] Field 4	
[05] Field 5	
[06] Field 6	
[07] Field 7	
[08] Field 8	
[09] Field 9	
[10] Field 10	
[11] Field 11	
[12] Field 12	
[13] Field 13	
[14] Field 14	
[15] Field 15	
[16] Field 16	
[17] Field 17	
[18] Field 18	
[19] Field 19	

ISOCENTER 0, 0, 0 (DICOM 15.5, 13.2, -18) 0, 0, 0 (DICOM 15.5, 13.2, -18)

MODALITY			ENERGY
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV
IMRT (Dynamic)	(166	CPs)	6FFF MV

GEOMETRY

ENERGY

Gantry: 189.5, Collimator: 270, Couch: 0 Gantry: 208.4, Collimator: 279.5, Couch: 0 Gantry: 227.4, Collimator: 289, Couch: 0 Gantry: 246.3, Collimator: 298.5, Couch: 0 Gantry: 265.3, Collimator: 308, Couch: 0 Gantry: 284.2, Collimator: 317.5, Couch: 0 Gantry: 303.2, Collimator: 327, Couch: 0 Gantry: 322.1, Collimator: 336.5, Couch: 0 Gantry: 341.1, Collimator: 346, Couch: 0 Gantry: 0, Collimator: 355.5, Couch: 0 Gantry: 19, Collimator: 5, Couch: 0 Gantry: 37.9, Collimator: 14.5, Couch: 0 Gantry: 56.9, Collimator: 24, Couch: 0 Gantry: 75.8, Collimator: 33.5, Couch: 0 Gantry: 94.8, Collimator: 43, Couch: 0 Gantry: 113.7, Collimator: 52.5, Couch: 0 Gantry: 132.7, Collimator: 62, Couch: 0 Gantry: 151.6, Collimator: 71.5, Couch: 0 Gantry: 170.6, Collimator: 81, Couch: 0

ProKnow®

METERSET	BEAM-ON TIME (Est.)
169.3988 MU	16
150.9321 MU	14
146.2013 MU	14
174.1715 MU	16
124.7384 MU	12
139.1884 MU	13
192.5639 MU	18
196.0265 MU	18
228.3189 MU	21
335.3096 MU	31
243.8767 MU	23
232.1199 MU	22
226.8647 MU	21
197.9261 MU	19
162.277 MU	15
213.2503 MU	20
131.1758 MU	12
171.7015 MU	16
151.8586 MU	14
	E EQ main (TOTAL)

3587.9 (TOTAL)

5.58 min (TOTAL)

MODIFIERS

X Jaws Y Jaws, MLC (X) X Jaws Y Jaws, MLC (X)



METRIC	RESULT	MIN REQ			
Volume (%) of the PTV63 covered by 63 (Gy)	98.794	90 🗹 ^{2.5}	p ls	5p 95	7.5p 97

Cumulative DVH: PTV63 (20.356 cc) Min: 61.033 Gy, Mean: 64.897 Gy, Max: 67.017 Gy,



Performance Bin Scoring

ProKnow®

0
_
_
_
_
_
_
_

55

50

60

65

70



METRIC	RESULT	MIN REQ		
Volume (%) of the PTV60 covered by 60 (Gy)	99.021	90	5p 95	7.5p 97

Cumulative DVH: PTV60 (90.026 cc) Min: 56.537 Gy, Mean: 62.389 Gy, Max: 65.654 Gy,



Performance Bin Scoring

ProKnow®

	IDEAL	PERFC	DRMANCE (F	PTS) WEIGHT
	108 108 100) (9.18)	10.00
Vol	: 90.026 c	C		

65

70


60

50

55




METRIC	RESULT	MIN REQ)				IDEAL	PERFO	RMANCE (PT	S) WEIGHT
Volume (%) of the PTV57 covered by 57 (Gy)	99.420	90	₹ 3.13p 90	6.25 95	p 9.38	0 12.5 100	100		(11.90)	12.50
Min: 43.7	Cum ′05 Gy, M∉	nulative ean: 59	e DVH: PTV .304 Gy, M	57 (431.3 lax: 68.27	77 cc) 7 Gy, \	/ol: 43	31.377 d	c		
100										
100										





METRIC	RESULT	MIN REQ IDEAL PERFORMANCE (PTS) WEIGH							
Volume (%) of the PTV54 covered by 54 (Gy)	99.121	90 😴	3.13p 90	6.25p 95	9.38p 97	12.5p 100	100	GOOD (11.58)	12.50
Min: 41.1	Cum 58 Gy, Me	ulative D ean: 55.7	VH: PTV54 (58 77 Gy, Max: 67.	6.877 523 G	cc) y, Vol	l: 58	6.877 cc		
100									






METRIC			RESULT MIN REQ					IDEAL	PERFORMANC	PERFORMANCE (PTS) WEIGHT			
ose (Gy) co	overing 99 (%	%) of the (CTV63	63.214	55	₹ 2.5p		59 59	7.5p 61	^{10p} 63	🗹 IDEAL (10.00)	10	0.00
									-				
			Min: 6	C 1.784 Gy	:umulat , Mean:	ive DVH: 65.226 (CTV63 Gy, Max	(11.07) : 67.017	l cc) ′ Gy, Vc	ol: 11.071	сс		
100													
90													
80													
70													
60													
50													
40													
30													
20													
10													
0	5	10	15	20	25	30	35	40	45	50	55 60	65	70







ProKnow®

		IDEAL		PERFORMANCE (PTS) WEIGHT						
.5p 9	10p 60	60		IDEAL	(10.00	D)	10.00			
		1								
Va	1. 5	8.950 c	~							
vu	1. 5	0.9000	<u> </u>		•					
					٦ _					
					+					
					\rightarrow					
						-\ -				




METRIC	RESULT	MIN REQ			
Dose (Gy) covering 99 (%) of the CTV57	57.606	54 🗸	2.5p 54	5p 55	7.5p 56





METRIC	RESULT MIN REQ					IDEAL	PERFORMANCE (P	TS) WEIGHT	
Dose (Gy) coverin	54.324	51	✓ 2.5p 51	5p 52	7.5p 53	^{10p} 54	🗹 IDEAL (10.00)	10.00	
	Min: 41.	Cur 084 Gy, M	mulativ 1ean: 5	e DVH: CT 5.846 Gy,	V54 (423 Max: 66.3	.925 cc) 74 Gy, Vo	l: 423.925	сс	
100									
90									
80									








	IDEAL	PERFO	PERFORMANCE (PTS) WEIGHT				
4p 5p 0.2 0.0	₃ 0.03		(4.63)		5.00		
Vol: 1	1.071 c	с					





Cumulative DVH: SpinalCord (33.780 cc Min: 2.391 Gy, Mean: 23.916 Gy, Max: 34.624 Gy, V




Performance Bin Scoring

	IDEAL			PERFORMANCE (PTS) WEIGHT					
	5p 30	30		GOOD (4.69)	5.00				
c) Vol: 33.780 cc									





Cumulative DVH: Brainstem (31.767 cc Min: 2.043 Gy, Mean: 12.232 Gy, Max: 32.386 Gy, V



Performance Bin Scoring

ProKnow®


	IDEAL	PERFORMANCE (PTS) WEIGHT					
5	8 30	GOOD	(4.90)	5.00			
c) Vol: 3	1.767 сс						

30

25








l	DEAL		PERF	RFORMANCE (PTS) WEIGHT						
5p 35	35		GOOI	D (4.32)		5.00				
с) Vol: 39.047 сс										
							1			






IDE/	AL F	PERFORMANC	CE (PTS) WEIGH	łT
4p 5p 15		GOOD (4.80)	5.00	
:) /ol: 34.54	3 cc			
				-
				-
				-
				-
				-
				-


40	45	50	55	60






	l	IDEAL		PERF	ORMANCE (PTS)	WEIGH	IT
	5p 26	26		GOOI	D (4.52)		5.00	
-)								
Vo	l: 39	.047	сс					
_								
Ц				L				





	IDE	AL	PER	FORM	ANCE	(PTS)	WEIGH	ΗT
	<u>₹</u> 8 20		GO	OD (4.	87)		5.00	
c) Vol: 3	34.54	l3 cc						
								_
								-
		" " U		lu, m	L I	. hu		
	U U			50		55		60
40		45		50		55		60







ProKnow[®]

	IDEAL	PERFO	ERFORMANCE (PTS) WEIGHT							
	§8 30	GOOD	(4.73)	5.00						
c) Vol: 65.720 cc										
4	կենքե	սավել է								











	IDEA	L	PER	FORM	ANCE	(PTS)	WEIGH	ΗT
5p 35	35		МА	RGINA	L (2.7	3)	5.00	
8.700 c Vol: 23	c) 3.70	0 cc						










Dose (Gy) --->

Performance Bin Scoring

	IDEAL	PE	RFORMANCE	E (PIS)	WEIGHT
2	58 60	<u></u> М	ARGINAL (2.	76)	5.00
c)	: 110.121	~~			
vor	. 110.121				
			<u> </u>		



METRIC	RESULT	MIN REQ		
Global maximum dose (Gy)	68.277	71 🗹 } 1	3p 69	4p 67



ProKnow

METRIC		RESULT	MIN REQ		
[CRITICAL] Number of u	nique isocenters	1.000	≤1	7 -10p	≤ 1
BEAM [#] NAME	MACHINE	MO	DALITY		ENERGY
[01] Field 1	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[02] Field 2	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[03] Field 3	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[04] Field 4	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[05] Field 5	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[06] Field 6	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[07] Field 7	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[08] Field 8	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[09] Field 9	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[10] Field 10	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[11] Field 11	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[12] Field 12	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[13] Field 13	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[14] Field 14	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[15] Field 15	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[16] Field 16	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[17] Field 17	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[18] Field 18	RDS3 Ti kV	IMRT	(Dynamic) (166	CPs)	6FFF MV
[19] Field 19	RDS3 Ti kV	IMRT	Oynamic) (166	CPs)	6FFF MV
BEAM [#] NAME	ISOCENTER			GEOMET	RY
[01] Field 1	0, 0, 0 (DICOM 15.5, 1	3.218)			89.5, Collimator: 270, Couch: 0
[02] Field 2	0, 0, 0 (DICOM 15.5, 1				08.4, Collimator: 279.5, Couch: 0
[03] Field 3	0, 0, 0 (DICOM 15.5, 1			-	27.4, Collimator: 289, Couch: 0
[04] Field 4	0, 0, 0 (DICOM 15.5, 1			-	46.3, Collimator: 298.5, Couch: 0
[05] Field 5	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		-	65.3, Collimator: 308, Couch: 0
[06] Field 6	0, 0, 0 (DICOM 15.5, 1			-	84.2, Collimator: 317.5, Couch: 0
[07] Field 7	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		-	03.2, Collimator: 327, Couch: 0
[08] Field 8	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 32	22.1, Collimator: 336.5, Couch: 0
[09] Field 9	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 34	41.1, Collimator: 346, Couch: 0
[10] Field 10	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 0,	Collimator: 355.5, Couch: 0
[11] Field 11	0, 0, 0 (DICOM 15.5, 1	.3.2, -18)		Gantry: 19	9, Collimator: 5, Couch: 0
[12] Field 12	0, 0, 0 (DICOM 15.5, 1	.3.2, -18)		Gantry: 37	7.9, Collimator: 14.5, Couch: 0
[13] Field 13	0, 0, 0 (DICOM 15.5, 1	.3.2, -18)		Gantry: 50	5.9, Collimator: 24, Couch: 0
[14] Field 14	0, 0, 0 (DICOM 15.5, 1	.3.2, -18)		Gantry: 75	5.8, Collimator: 33.5, Couch: 0
[15] Field 15	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 94	4.8, Collimator: 43, Couch: 0
[16] Field 16	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 11	13.7, Collimator: 52.5, Couch: 0
[17] Field 17	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 13	32.7, Collimator: 62, Couch: 0
[18] Field 18	0, 0, 0 (DICOM 15.5, 1	3.2, -18)		Gantry: 15	51.6, Collimator: 71.5, Couch: 0
[19] Field 19	0, 0, 0 (DICOM 15.5, 1	.3.2, -18)		Gantry: 17	70.6, Collimator: 81, Couch: 0

Performance Bin Scoring

ProKnow®

IDEAL	PERFORMANCE (PTS) WEIGHT
Ор	🗹 IDEAL (0.00)	0.00
METERSET	BEAM-ON TIME	(Est.)
169.3988 MU	16	
150.9321 MU	14	
146.2013 MU	14	
174.1715 MU	16	
124.7384 MU	12	
139.1884 MU	13	
192.5639 MU	18	
196.0265 MU	18	
228.3189 MU	21	
335.3096 MU	31	
243.8767 MU	23	
232.1199 MU	22	
226.8647 MU	21	
197.9261 MU	19	
162.277 MU	15	
213.2503 MU	20	
131.1758 MU	12	
171.7015 MU	16	
151.8586 MU	14	
3587.9 (TOTAL)	5.58 min (TOTAL	.)

MODIFIERS

X Jaws Y Jaws, MLC (X)
X Jaws Y Jaws, MLC (X)

ProKnow

METRIC		RESULT	MIN REQ		
[CRITICAL] Number of u	inique couch angles	1.000	≤1 🗸	- 10p	≤ 1
BEAM [#] NAME	MACHINE	мо	DALITY	E	NERGY
[01] Field 1	RDS3 Ti kV	IMRT	(Dynamic) (166 C	(Ps) 6	FFF MV
[02] Field 2	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[03] Field 3	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[04] Field 4	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[05] Field 5	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[06] Field 6	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[07] Field 7	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[08] Field 8	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[09] Field 9	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[10] Field 10	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[11] Field 11	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[12] Field 12	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[13] Field 13	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[14] Field 14	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[15] Field 15	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[16] Field 16	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[17] Field 17	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[18] Field 18	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
[19] Field 19	RDS3 Ti kV		(Dynamic) (166 C		FFF MV
BEAM [#] NAME	ISOCENTER			GEOMETRY	
[01] Field 1	0, 0, 0 (DICOM 15.5, 13.	218)			5, Collimator: 270, Couch: 0
[02] Field 2	0, 0, 0 (DICOM 15.5, 13.				4, Collimator: 279.5, Couch: 0
[03] Field 3	0, 0, 0 (DICOM 15.5, 13.				4, Collimator: 289, Couch: 0
[04] Field 4	0, 0, 0 (DICOM 15.5, 13.				3, Collimator: 298.5, Couch: 0
[05] Field 5	0, 0, 0 (DICOM 15.5, 13.				3, Collimator: 308, Couch: 0
[06] Field 6	0, 0, 0 (DICOM 15.5, 13.				2, Collimator: 317.5, Couch: 0
[07] Field 7	0, 0, 0 (DICOM 15.5, 13.			2	2, Collimator: 327, Couch: 0
[08] Field 8	0, 0, 0 (DICOM 15.5, 13.				1, Collimator: 336.5, Couch: 0
[09] Field 9	0, 0, 0 (DICOM 15.5, 13.			2	1, Collimator: 346, Couch: 0
[10] Field 10	0, 0, 0 (DICOM 15.5, 13.				ollimator: 355.5, Couch: 0
[11] Field 11	0, 0, 0 (DICOM 15.5, 13.				Collimator: 5, Couch: 0
[12] Field 12	0, 0, 0 (DICOM 15.5, 13.			2	, Collimator: 14.5, Couch: 0
[13] Field 13	0, 0, 0 (DICOM 15.5, 13.			-	, Collimator: 24, Couch: 0
[14] Field 14	0, 0, 0 (DICOM 15.5, 13.			-	, Collimator: 33.5, Couch: 0
[15] Field 15	0, 0, 0 (DICOM 15.5, 13.			-	, Collimator: 43, Couch: 0
[16] Field 16	0, 0, 0 (DICOM 15.5, 13.			-	7, Collimator: 52.5, Couch: 0
[17] Field 17	0, 0, 0 (DICOM 15.5, 13.				7, Collimator: 62, Couch: 0
[18] Field 18	0, 0, 0 (DICOM 15.5, 13.			2	6, Collimator: 71.5, Couch: 0
[19] Field 19	0, 0, 0 (DICOM 15.5, 13.				6, Collimator: 81, Couch: 0
	· · · · · · · · · · · · · · · · · · ·			,	

Performance Bin Scoring

ProKnow®

IDEAL	PERFORMANCE (PTS) WEIGHT
Ор	🗹 IDEAL (0.00)	0.00
METERSET	BEAM-ON TIME	(Est.)
169.3988 MU	16	
150.9321 MU	14	
146.2013 MU	14	
174.1715 MU	16	
124.7384 MU	12	
139.1884 MU	13	
192.5639 MU	18	
196.0265 MU	18	
228.3189 MU	21	
335.3096 MU	31	
243.8767 MU	23	
232.1199 MU	22	
226.8647 MU	21	
197.9261 MU	19	
162.277 MU	15	
213.2503 MU	20	
131.1758 MU	12	
171.7015 MU	16	
151.8586 MU	14	
3587.9 (TOTAL)	5.58 min (TOTAL	.)

MODIFIERS

X Jaws Y Jaws, MLC (X)
X Jaws Y Jaws, MLC (X)



METRIC		RESULT	MIN REC	Ś		
Cumulative meterset o	ver all treatment beams	3587.900				
BEAM [#] NAME	MACHINE	MOD	ALITY		ENERGY	
[01] Field 1	RDS3 Ti kV	IMRT	(Dynamic) (1	66 CPs)	6FFF MV	
[02] Field 2	RDS3 Ti kV	IMRT	(Dynamic) (1	.66 CPs)	6FFF MV	
[03] Field 3	RDS3 Ti kV	IMRT	(Dynamic) (1	66 CPs)	6FFF MV	

[03] Field 3	RDS3 TI KV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[04] Field 4	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[05] Field 5	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[06] Field 6	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[07] Field 7	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[08] Field 8	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[09] Field 9	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[10] Field 10	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[11] Field 11	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[12] Field 12	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[13] Field 13	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[14] Field 14	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[15] Field 15	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[16] Field 16	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[17] Field 17	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[18] Field 18	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	
[19] Field 19	RDS3 Ti kV	IMRT (Dynamic) (166 CPs)	6FFF MV	

GEOMETRY

Gantry: 189.5, Collimator: 270, Couch: 0

Gantry: 227.4, Collimator: 289, Couch: 0

Gantry: 265.3, Collimator: 308, Couch: 0

Gantry: 303.2, Collimator: 327, Couch: 0

Gantry: 341.1, Collimator: 346, Couch: 0

Gantry: 37.9, Collimator: 14.5, Couch: 0

Gantry: 75.8, Collimator: 33.5, Couch: 0

Gantry: 113.7, Collimator: 52.5, Couch: 0

Gantry: 151.6, Collimator: 71.5, Couch: 0

Gantry: 132.7, Collimator: 62, Couch: 0

Gantry: 170.6, Collimator: 81, Couch: 0

Gantry: 56.9, Collimator: 24, Couch: 0

Gantry: 94.8, Collimator: 43, Couch: 0

Gantry: 0, Collimator: 355.5, Couch: 0

Gantry: 19, Collimator: 5, Couch: 0

Gantry: 208.4, Collimator: 279.5, Couch: 0

Gantry: 246.3, Collimator: 298.5, Couch: 0

Gantry: 284.2, Collimator: 317.5, Couch: 0

Gantry: 322.1, Collimator: 336.5, Couch: 0

BEAM	[#]	NAME
------	-----	------

BEAM [#] NAME	ISOCENTER
[01] Field 1	0, 0, 0 (DICOM 15.5, 13.2, -18)
[02] Field 2	0, 0, 0 (DICOM 15.5, 13.2, -18)
[03] Field 3	0, 0, 0 (DICOM 15.5, 13.2, -18)
[04] Field 4	0, 0, 0 (DICOM 15.5, 13.2, -18)
[05] Field 5	0, 0, 0 (DICOM 15.5, 13.2, -18)
[06] Field 6	0, 0, 0 (DICOM 15.5, 13.2, -18)
[07] Field 7	0, 0, 0 (DICOM 15.5, 13.2, -18)
[08] Field 8	0, 0, 0 (DICOM 15.5, 13.2, -18)
[09] Field 9	0, 0, 0 (DICOM 15.5, 13.2, -18)
[10] Field 10	0, 0, 0 (DICOM 15.5, 13.2, -18)
[11] Field 11	0, 0, 0 (DICOM 15.5, 13.2, -18)
[12] Field 12	0, 0, 0 (DICOM 15.5, 13.2, -18)
[13] Field 13	0, 0, 0 (DICOM 15.5, 13.2, -18)
[14] Field 14	0, 0, 0 (DICOM 15.5, 13.2, -18)
[15] Field 15	0, 0, 0 (DICOM 15.5, 13.2, -18)
[16] Field 16	0, 0, 0 (DICOM 15.5, 13.2, -18)
[17] Field 17	0, 0, 0 (DICOM 15.5, 13.2, -18)
[18] Field 18	0, 0, 0 (DICOM 15.5, 13.2, -18)
[19] Field 19	0, 0, 0 (DICOM 15.5, 13.2, -18)

Performance Bin Scoring ProKnow®

IDEAL

PERFORMANCE (PTS) WEIGHT

METERSET
169.3988 MU
150.9321 MU
146.2013 MU
174.1715 MU
124.7384 MU
139.1884 MU
192.5639 MU
196.0265 MU
228.3189 MU
335.3096 MU
243.8767 MU
232.1199 MU
226.8647 MU
197.9261 MU
162.277 MU
213.2503 MU
131.1758 MU
171.7015 MU
151.8586 MU
3587.9 (TOTAL)

BEAM-ON TIME (Est.)

16 14

- 14
- 16 12
- 13
- 18
- 18
- 21
- 31
- 23
- 22
- 21
- 19
- 15 20

12 16 14

- 5.58 min (TOTAL)

MODIFIERS

X Jaws Y Jaws, MLC (X) X Jaws Y Jaws, MLC (X)